

Figure 1. Stored electrogram




Figure 2. Stored electrograms

Consider these cases. What happened and what reasons led to this?

- A. Inappropriate ICD discharges: noise on the RV electrode - Fig.1, Rapid atrial fibrillation - Fig.2
- B. Inappropriate ICD discharges: T wave oversensing - Fig.1, Rapid atrial flutter - Fig.2
- C. Appropriate ICD discharges: Ventricular fibrillation - Fig.1, Rapid ventricular tachycardia - Fig.2
- D. Inappropriate ICD discharges: T wave oversensing - Fig.1, Pacemaker-mediated tachycardia - Fig.2

Ayan Abdrakhmanov, Zhandos Esilbayev
Department of Interventional Arrhythmology
National Research Cardiac Surgery Center
Nur-Sultan, Kazakhstan

Peer-review: Internal

Conflict of interest: None to declare

Authorship: A.A. and Z.E. equally contributed to preparation of quiz

Acknowledgement and funding: None to declare

Address for Correspondence: Ayan Abdrakhmanov, Department of Interventional Arrhythmology, National Research Cardiac Surgery Center, Nur-Sultan, Kazakhstan, Email: ayan-3@mail.ru

Received: 24.01.2019 **Accepted:** 25.01.2019

Copyright © 2019 Heart, Vessels and Transplantation