

Answer to Quiz on page... : ICD

**Correct Answer A: Inappropriate ICD discharges: noise on the RV electrode (Fig.1)
Rapid atrial fibrillation (Fig.2)**

Ayan Abdrakhmanov, Zhandos Esilbayev
Department of Interventional
Arrhythmology
National Research Cardiac Surgery
Center
Nur-Sultan, Kazakhstan

Peer-review: Internal

Conflict of interest: None to declare

Authorship: A.A. and Z.E. equally
contributed to preparation of quiz

Acknowledgement and funding: None to
declare

Address for Correspondence: Ayan Abdrakhmanov, Department of Interventional Arrhythmology,
National Research Cardiac Surgery Center, Nur-Sultan, Kazakhstan, Email: ayan-3@mail.ru

Received 24.01.2019 **Accepted:** 25.01.2019

Copyright © 2019 Heart, Vessels and Transplantation

doi: 10.24969/hvt.2019.146